UNITED STATES PATENT AND TRADEMARK OFFICE

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/565,863	01/24/2006	Yoshitomo Takaishi	D3301- 00160	7327

8933   7590   06/24/2010
DUANE MORRIS LLP - Philadelphia
IP DEPARTMENT
30 SOUTH 17TH STREET
PHILADELPHIA, PA 19103-4196

EXAMINER
LEE, JOHN W

ART UNIT	PAPER NUMBER
2624	

MAIL DATE	DELIVERY MODE
06/24/2010	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

PTOL-90A (Rev. 04/07)

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE *3* MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.
- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

1)☒ Responsive to communication(s) filed on *22 February 2010*.
2a)☐ This action is **FINAL**.　　　　2b)☒ This action is non-final.
3)☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

4)☒ Claim(s) *1-17* is/are pending in the application.
　　4a) Of the above claim(s) *2-5 and 11* is/are withdrawn from consideration.
5)☐ Claim(s) _____ is/are allowed.
6)☒ Claim(s) *1,6-10 and 12-17* is/are rejected.
7)☐ Claim(s) _____ is/are objected to.
8)☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

9)☒ The specification is objected to by the Examiner.
10)☒ The drawing(s) filed on *24 January 2006* is/are: a)☒ accepted or b)☐ objected to by the Examiner.
　　Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
　　Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
11)☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

12)☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
　　a)☒ All　b)☐ Some * c)☐ None of:
　　　1.☒ Certified copies of the priority documents have been received.
　　　2.☐ Certified copies of the priority documents have been received in Application No. _____.
　　　3.☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
　　* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

1)☒ Notice of References Cited (PTO-892)
2)☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3)☐ Information Disclosure Statement(s) (PTO/SB/08)
　　Paper No(s)/Mail Date _____.
4)☐ Interview Summary (PTO-413)
　　Paper No(s)/Mail Date. _____.
5)☐ Notice of Informal Patent Application
6)☐ Other: _____.

DETAILED ACTION

- Claims 1, 6-10 and 12-17 are pending; claims 2-5 and 11 are canceled; claims 1, 12 and 14 are amended.

- The application has entered the national stage from an international application, PCT/JP04/10815, after compliance with 35 U.S.C. § 371.

Continued Examination Under 37 CFR 1.114

1. A request for continued examination under 37 CFR 1.114, including the fee set forth in 37 CFR 1.17(e), was filed in this application after final rejection. Since this application is eligible for continued examination under 37 CFR 1.114, and the fee set forth in 37 CFR 1.17(e) has been timely paid, the finality of the previous Office action has been withdrawn pursuant to 37 CFR 1.114. Applicant's submission filed on 22 February 2010 has been entered.

Response to Amendments/Arguments

2. Applicant's amendment filed 18 May 2010 with respect to claims 1, 6-10 and 12-17 are entered and have been fully considered.

3.   Applicant's arguments of the rejected claims 1, 6-10, 12-17 under 35 U.S.C. §

103(a)[1] have been considered but are moot in view of the new ground(s) of rejection.

Priority

4.   Acknowledgment is made of applicant's claim for foreign priority under 35

U.S.C. 119(a)-(d). The certified copies have been filed in parent Applications No. JP

2003-283686 and JP 2004-185372, filed on 2003 July 31 and 2004 June 23,

respectively.

5.   Receipt is acknowledged of papers submitted under 35 U.S.C. 119(a)-(d), which

papers have been placed of record in the file.

Specification

6.   Applicant is reminded of the proper language and format for an abstract of the

disclosure:

> The abstract should be in narrative form and generally limited to a single
> paragraph on a separate sheet **within the range of 50 to 150 words** [emphasis
> added]. It is important that the abstract not exceed 150 words in length since the
> space provided for the abstract on the computer tape used by the printer is
> limited. The form and legal phraseology often used in patent claims, such as
> "means" and "said," should be avoided. The abstract should describe the
> disclosure sufficiently to assist readers in deciding whether there is a need for
> consulting the full patent text for details.

---

[1] Rejection of claims 1, 6-7, 9-10, 12-14 and 16-17 under 35 U.S.C. § 103(a) as being unpatentable over Lang et al. (WO 02/30283

A2); rejection of claims 8 and 15 under 35 U.S.C. § 103(a) as being unpatentable over Lang et al. (WO 02/30283 A2) in view of

Inoue (US 6,819,794 B2)

The language should be clear and concise and should not repeat information given in the title. It should avoid using phrases which can be implied, such as, "The disclosure concerns," "The disclosure defined by this invention," "The disclosure describes," etc[2].

a.   The abstract has 166 words which is more than 150 words.

Claim Rejections - 35 USC § 112

7.   The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

8.   Claim 13 is rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention.

a.   Claim 13 have insufficient antecedent basis for the limitation in the claim.

Regarding claim 13, "said standard value" (line 2) has a lack of antecedent basis because the claim contains no earlier recitation or limitation of "a standard value," and it would be unclear as to what element "said standard value" is making reference.

Claim Rejections - 35 USC § 103

9.   The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

---

[2] See MPEP § 608.01(b)

10.   Claims 1 and 6-10 are rejected under 35 U.S.C. 103(a) as being unpatentable

over Choel et al. ("Trabecular alveolar bone in the human mandible: A dual energy x-ray

absorptiometry study") in view of Inoue (US 2001/0021269).

        a.   Regarding claim 1, Choel discloses a bone mineral density evaluation

system for evaluating a bone mineral density from an X-ray picture (Objective; page 364,

"evaluate the potential use of dual energy x-ray absorptiometry for the assessment of

bone mineral content and bone mineral density") of a mandible (Fig. 1; Chapter-Material

and Methods: Ex vivo materials, page 365, "mandibular bone"), said X-ray picture

containing a picture of a specimen (Fig. 2; Chapter-Material and Methods: Ex vivo

materials, page 365, "The author discloses that three regions of interest- G, R1 and R2

were delineated from the three specimens- Incisal specimen (IS), premolar specimen

(PS) and molar specimen (MS), by the x-ray.") disposed beside said mandible (Fig. 1;

Chapter-Material and Methods: Ex vivo materials, page 365, "mandibular bone"), said

system comprising:

        detecting means for detecting a gradation particular portion of said picture of said

specimen (Fig. 2; Chapter-Material and Methods: Ex vivo materials, page 365, "The

author discloses that three regions of interest- G, R1 and R2 were delineated from the

three specimens- Incisal specimen (IS), premolar specimen (PS) and molar specimen

(MS), by the x-ray. Moreover, x-ray pictures have gradation level- black and white.");

and

        evaluating means for evaluating the bone mineral density means (Table III;

Chapter- Material and Methods: Ex vivo materials and Chatper-BMD differences related

to sex, dental status, and anatomic location, pages 365 and 367, "The author discloses using a multiple regression analysis based on the mean bone mineral density (BMD) and SDs for G, R1 and R2, which are the global specimen, infra-alveolar ROI and real intra-alveolar ROI, respectively, to assess the sex and dental status.") on the basis of the gradation of said X-ray picture (Fig. 2; Chapter-Material and Methods: Ex vivo materials, page 365, "The author discloses that three regions of interest- G, R1 and R2 were delineated from the three specimens- Incisal specimen (IS), premolar specimen (PS) and molar specimen (MS), by the x-ray. Moreover, x-ray pictures have gradation level- black and white.")

wherein:

said evaluating means makes evaluation (Table III; Chapter- Material and Methods: Ex vivo materials and Chatper-BMD differences related to sex, dental status, and anatomic location, pages 365 and 367, "The author discloses using a multiple regression analysis based on the mean bone mineral density (BMD) and SDs for G, R1 and R2, which are the global specimen, infra-alveolar ROI and real intra-alveolar ROI, respectively, to assess the sex and dental status.") on the basis of the gradation of a particular region (Fig. 2; Chapter-Material and Methods: Ex vivo materials, page 365, "The author discloses that three regions of interest- G, R1 and R2 were delineated from the three specimens- Incisal specimen (IS), premolar specimen (PS) and molar specimen (MS), by the x-ray.") of said mandible (Fig. 1; Chapter-Material and Methods: Ex vivo materials, page 365, "mandibular bone") in said X-ray picture (Objective; page 364, "dual energy x-ray absorptiometry"); and

said particular region (Fig. 2; Chapter-Material and Methods: Ex vivo materials,

page 365, "The author discloses three regions of interest- G, R1 and R2.") includes a

region corresponding to an alveolar bone portion (Chapter- Material and Methods: Ex

vivo materials, page 365, "The author discloses R1 and R2 are infra-alveolar ROI and

real intra-alveolar ROI, respectively.") around a first premolar (Fig. 1; Chapter- Material

and Methods: Ex vivo materials, page 365, "premolar specimen (PS)").

However, Choel does not disclose correcting means for correcting the gradation

of said X-ray picture so as to make the gradation of said particular portion of said picture

of said specimen as detected by said detecting means comply with a standard value to

which said gradation of said particular portion is referenced.

Instead of Choel, Inoue, the same field of endeavor of medical digital x-ray image

processing, discloses correcting means for correcting (Fig. 1; "convert an input image

from which the histogram 101 is obtained to an aimed image (image in an ideal state)

for which the histogram 102 is obtained" at ¶¶ 0052 and 0057) the gradation of said X-

ray picture (Fig. 1-101, "histogram of input image"; "histogram of an X-ray dose (pixel

value) of an image (input image) of a specific field of a subject" at ¶0047) so as to make

the gradation of said particular portion of said picture of said specimen (Fig. 1-101,

"histogram of input image"; "histogram of an X-ray dose (pixel value) of an image (input

image) of a specific field of a subject" at ¶0047) as detected by said detecting means

comply (Fig. 1-106, "broadly equalized histogram"; "[G]enerally equalized histogram is

captuered" at ¶¶ 0050 and 0051) with a standard value (Fig. 1-102, "aimed histogram";

a histogram (aimed histogram) of an X-ray dose (pixel value) of an image in an ideal

state (hereinafter referred to as "an aimed image") of a specific field (identical with a

field in an input image) of a subject" at ¶ 0051) to which said gradation of said particular

portion (Fig. 1-101, "histogram of input image"; "histogram of an X-ray dose (pixel

value) of an image (input image) of a specific field of a subject" at ¶0047) is referenced.

Choel and Inoue are combinable because they are all related to the field of

medical digital x-ray image processing.

At the time of the invention, it would have been obvious to a person of ordinary

skill in the art to applying the steps of detecting the "histogram of an X-ray dose (pixel

value) of an image (input image) of a specific field of a subject" (Inoue; Fig. 101; ¶ 0047)

and "converting an input image[,] from which the histogram 101 is obtained [,] to an

aimed image (image in an ideal state)[,] for which the histogram 102 is obtained" (Inoue;

Fig. 1; ¶¶ 0052 and 0057), using "the generally equalized histogram [being] captured"

(Inoue; Fig. 1-106; ¶ 0050 and 0051) taught by Inoue in the process of delineating the

"three region (ROIs: G, R1, R2)" (Choel; Fig. 2; Chapter Material and Methods: Ex vivo

materials, page 365) of Choel's method of "evaluate the potential use of dual energy x-

ray absorptiometry for the assessment of bone mineral content and bone mineral

density" (Choel; Objective; page 364).

The suggestion/motivation for doing so would have been to eliminate

"unstableness of gradation conversion processing [of] a complicated operation such as

analysis of an object image itself analysis of a histogram of an object image intervenes

of the like, [which] a lot of time is required for analysis ,computational processing or the

like, and in some object images, an analysis mistake may occur" (Inoue; ¶ 0025) by

providing "a [process] of changing a gradation conversion characteristics used in

applying gradation conversion processing to an aimed image" (Inoue; ¶ 0046) and

"utilize[ing] equalization of a histogram" (Inoue; ¶ 0046) that will "enable realization of

stable gradation conversion easily and efficiently" (Inoue; ¶ 0027).

Therefore, it would have been obvious to combine Choel and Inoue to obtain the

invention as recited in claim 1.

b.   Regarding claim 6, the combination of Choel and Inoue, as applied in

claim 1, discloses all the previous claim limitations. Moreover, Inoue discloses further

comprising setting means for setting said standard value ("setting a form of an ideal

histogram" at ¶ 0067).

Choel and Inoue are combinable because they are all related to the field of

medical digital x-ray image processing.

At the time of the invention, it would have been obvious to a person of ordinary

skill in the art to apply "setting a form of an ideal histogram" (Inoue; ¶ 0067) taught by

Inoue in the process of delineating the "three region (ROIs: G, R1, R2)" (Choel; Fig. 2;

Chapter Material and Methods: Ex vivo materials, page 365) of the combination of

Choel and Inoue.

The suggestion/motivation for doing so would have been to enable "a user [] to

select an aimed field … and to set parameters for image processing with respect to the

aimed field" (Inoue; ¶ 0075) because "a method of image processing or its conditions is

different for each field of a human body that requires treatments or diagnosis in the [] X-

ray photographing apparatus or system [] in order to perform image processing peculiar

to each field" (Inoue; ¶ 0075).

Therefore, it would have been obvious to combine Choel and Inoue to obtain the

invention as specified in claim 6.

c.   Regarding claim 7, the combination of Choel and Inoue, as applied in

claim 1, discloses wherein said standard value (Inoue; Fig. 1-102, "aimed histogram"; a

histogram (aimed histogram) of an X-ray dose (pixel value) of an image in an ideal state

(hereinafter referred to as "an aimed image") of a specific field (identical with a field in

an input image) of a subject" at ¶ 0051) being set based on a result of detection by said

detecting means of a particular X-ray picture (Inoue; "The author discloses the aimed

histogram is a specific field, which is identical with a field in an input image that is

captured by X-ray photographing, of a subject. The specific field of the subject, which is

a human body disclosed in Inoue, is the result of detection by the detecting means of a

X-ray photography" ¶ 0047 and 0051).

d.   Regarding claim 8, the combination of Choel and Inoue, as applied in

claim 1, discloses all the previous claim limitation including wherein said evaluating

means (Choel;Table III; Chapter- Material and Methods: Ex vivo materials and Chatper-

BMD differences related to sex, dental status, and anatomic location, pages 365 and

367, "The author discloses using a multiple regression analysis based on the mean

bone mineral density (BMD) and SDs for G, R1 and R2, which are the global specimen,

infra-alveolar ROI and real intra-alveolar ROI, respectively, to assess the sex and dental

status."). Moreover, Inoue discloses display means (Fig. 3-310; "display" at ¶ 0118) for

displaying said corrected gradation in the form of histogram (Fig. 3-307; "The author

discloses that a histogram, which is created after gradation conversion has passed

through the memory, is displayed on the display as the histogram" at ¶ 0118).

Choel and Inoue are combinable because they are all related to the field of

medical digital x-ray image processing.

At the time of the invention, it would have been obvious to a person of ordinary

skill in the art to apply the step of "[t]he created histogram [being] displayed on the

display 310 as the histogram 307" (Inoue; Figs. 3-307 and 3-310; ¶ 0118) in the step of

the "multiple regression analysis … [which] sex and dental status [being] assessed'

(Chole; Table III; Chapter- Material and Methods: Ex vivo materials and Chatper-BMD

differences related to sex, dental status, and anatomic location, pages 365 and 367) of

the combination of Chole and Inoue.

The suggestion/motivation for doing so would have been to provide "[t]he created histogram [being] displayed on the display at the histogram" (Inoue; ¶ 0118) because "an observer can observe most easily" (Inoue; ¶ 0067) "a histogram of an image ... while gradationally converting one or a plurality of images interactively" (Inoue; ¶ 0067).

Therefore, it would have been obvious to combine Choel and Inoue to obtain the invention as specified in claim 8.

e.   Regarding claim 9, the combination of Choel and Inoue, as applied in claim 1, discloses wherein said evaluating means includes judging means for judging said bone mineral density (Choel; Fig. 2; Table III; Chapter-BMD differences related to sex, dental status, and anatomic location, page 367, "The author discloses using a multiple regression analysis based on the mean bone mineral density (BMD) and SDs for G, R1 and R2 to assess the sex and dental status.") on the basis of said corrected gradation (Inoue; Fig. 1; "convert an input image from which the histogram 101 is obtained to an aimed image (image in an ideal state) for which the histogram 102 is obtained" at ¶¶ 0052 and 0057).

f.   Regarding claim 10, the combination of Choel and Inoue, as applied in claim 1, discloses further comprising output means for providing together a plurality of evaluation results provided by said evaluation means (Choel; Fig. 2; Table III; Chapter-BMD differences related to sex, dental status, and anatomic location, page 367, "The author discloses using a multiple regression analysis based on the mean bone mineral

density (BMD) and SDs for G, R1 and R2 to assess the sex and dental status.") for

respective ones of a plurality of X-ray pictures (Choel; Fig. 2; Table III; Chapter-BMD

differences related to sex, dental status, and anatomic location, page 367, "G, R1 and

R2 obtained for male and female specimens, and dentate and edentulous specimens").

11.   Claims 12-14 and 16-17 are rejected under 35 U.S.C. 103(a) as being

unpatentable over Choel et al. ("Trabecular alveolar bone in the human mandible: A

dual energy x-ray absorptiometry study") in view of Kim (US 6,078,686).

         a.   Regarding claim 12, Choel discloses a bone mineral density evaluation

system for evaluating a bone mineral density from an X-ray picture (Objective; page 364,

"evaluate the potential use of dual energy x-ray absorptiometry for the assessment of

bone mineral content and bone mineral density") of a mandible (Fig. 1; Chapter-Material

and Methods: Ex vivo materials, page 365, "mandibular bone"), said X-ray picture

containing a picture of a specimen (Fig. 2; Chapter-Material and Methods: Ex vivo

materials, page 365, "The author discloses that three regions of interest- G, R1 and R2

were delineated from the three specimens- Incisal specimen (IS), premolar specimen

(PS) and molar specimen (MS), by the x-ray.") disposed beside said mandible (Fig. 1;

Chapter-Material and Methods: Ex vivo materials, page 365, "mandibular bone"), a

gradation of said picture of said specimen varying from portion to portion (Fig. 2;

Chapter-Material and Methods: Ex vivo materials, page 365, "The author discloses that

three regions of interest- G, R1 and R2 were delineated from the three specimens-

Incisal specimen (IS), premolar specimen (PS) and molar specimen (MS), by the x-ray.

Moreover, x-ray pictures have gradation level- black and white."), said system

comprising:

detecting means for detecting the gradation of said picture of said specimen (Fig.

2; Chapter-Material and Methods: Ex vivo materials, page 365, "The author discloses

that three regions of interest- G, R1 and R2 were delineated from the three specimens-

Incisal specimen (IS), premolar specimen (PS) and molar specimen (MS), by the x-ray.

Moreover, x-ray pictures have gradation level- black and white."); and

evaluating means for evaluating the bone mineral density (Table III; Chapter-

Material and Methods: Ex vivo materials and Chatper-BMD differences related to sex,

dental status, and anatomic location, pages 365 and 367, "The author discloses using a

multiple regression analysis based on the mean bone mineral density (BMD) and SDs

for G, R1 and R2, which are the global specimen, infra-alveolar ROI and real intra-

alveolar ROI, respectively, to assess the sex and dental status.") on the basis of the

gradation of said X-ray picture (Fig. 2; Chapter-Material and Methods: Ex vivo materials,

page 365, "The author discloses that three regions of interest- G, R1 and R2 were

delineated from the three specimens- Incisal specimen (IS), premolar specimen (PS)

and molar specimen (MS), by the x-ray. Moreover, x-ray pictures have gradation level-

black and white."),

wherein said evaluating means makes evaluation (Table III; Chapter- Material

and Methods: Ex vivo materials and Chatper-BMD differences related to sex, dental

status, and anatomic location, pages 365 and 367, "The author discloses using a

multiple regression analysis based on the mean bone mineral density (BMD) and SDs

for G, R1 and R2, which are the global specimen, infra-alveolar ROI and real intra-alveolar ROI, respectively, to assess the sex and dental status.") on the basis of the gradation of a particular region of said mandible in said X-ray picture (Fig. 2; Chapter-Material and Methods: Ex vivo materials, page 365, "The author discloses that three regions of interest- G, R1 and R2 were delineated from the three specimens- Incisal specimen (IS), premolar specimen (PS) and molar specimen (MS), by the x-ray. Moreover, x-ray pictures have gradation level- black and white."),

wherein said particular region (Fig. 2; Chapter-Material and Methods: Ex vivo materials, page 365, "The author discloses three regions of interest- G, R1 and R2.") includes a region corresponding to an alveolar bone portion (Chapter- Material and Methods: Ex vivo materials, page 365, "The author discloses R1 and R2 are infra-alveolar ROI and real intra-alveolar ROI, respectively.") around a first premolar (Fig. 1; Chapter- Material and Methods: Ex vivo materials, page 365, "premolar specimen (PS)").

However, Choel does not disclose detecting means for detecting an average and a deviation of the gradation of said picture and correcting means for correcting the gradation of said X-ray picture so as to make the average and the deviation as detected by said detecting means comply with a standard average and a standard deviation to which said detected average and said detected deviation are referenced, respectively.

Instead of Choel, Kim, the same field of endeavor of image processing, discloses detecting means for detecting an average (FIG. 4-304, "frame mean calculator; "A frame mean calculator calculates the mean level (X_m)" at col. 8, lines 38-39) and a deviation (FIGs. 4-308, "a first CDF calculator" and 4-310, "a second CDF calculator"; equations

(15) and (16); "calculate[ing] a cumulative density function (CDF) $c_L(X_k)$ …

calculate[ing] a cumulative density function (CDF) $c_U(X_k)$" at col. 8, lines 63-67 to col. 9,

lines 1-18) of the gradation of said picture ("[A]n input image signal {Y} [being]

comprised of L discrete level represented by {$\{X_0, X_1, ..., X_{L-1}\}$" at col. 8, lines 36-37)

and

correcting means for correcting the gradation (equation (18); "equalized output

(Yo)" at col. 10, lines 22-25 and 37-38) of said picture ("[A]n input image signal {Y}

[being] comprised of L discrete level represented by {$\{X_0, X_1, ..., X_{L-1}\}$" at col. 8, lines

36-37) so as to make the average (FIG. 4-304, "frame mean calculator; "A frame mean

calculator calculates the mean level (X_m)" at col. 8, lines 38-39) and the deviation

(FIGs. 4-308, "a first CDF calculator" and 4-310, "a second CDF calculator"; equations

(15) and (16); "calculate[ing] a cumulative density function (CDF) $c_L(X_k)$ …

calculate[ing] a cumulative density function (CDF) $c_U(X_k)$" at col. 8, lines 63-67 to col. 9,

lines 1-18) as detected by said detecting means comply with a standard average

(equation (17); "a compensated mean level (B_m) $B_m = X_m + \Delta$" at col. 9, lines 34-35)

and a standard deviation (equations (18) and (19); "$c_L(X_k)B_m$" and

"$B'_m + (X_{L-1} - B'_m)c_U(X_k)$" at col. 10, lines 21-25) to which said detected average and

said detected deviation are referenced, respectively (FIG. 4-314, "brightness

compensator"; "The inventor discloses that a corrected value (Δ) was added to the

mean level (X_m) to obtain a compensated mean level (B_m) for the adjustment or

compliance" at col. 9, lines 29-67; FIGS. 4-316, "the first mapper" and 4-318, "the

second mapper"; "The inventor discloses that the first mapper calculates $\,^{\prime\prime}c_L(X_k)B_m\,^{\prime\prime}$

using a cumulative density function (CDF) $c_L(X_k)$ and the second mapper calculates

$\,^{\prime\prime}B'_m+(X_{L-1}-B'_m)c_U(X_k)\,^{\prime\prime}$ using a cumulative density function (CDF) $c_U(X_k)$ for the

adjustment or compliance" at col. 10, lines 5-36).

Choel and Kim are combinable because they are all related to the field of image

processing.

At the time of the invention, it would have been obvious to a person of ordinary

skill in the art to apply the "contrast enhancer" (KIM; FIG. 1-300; col. 8, lines 29-30)

comprising "the frame mean calculator" (Kim; FIG. 4-304; col. 8, lines 38-39), "the first

and second CDF calculator (Kim; equations (15) and (16); FIGS. 4-308 and 4-310; col.

8, lines 63-67 to col. 9, lines 1-18), "brightness compensator" (Kim; FIG. 4-314; col. 9,

lines 29-67) and "the first and second mapper" (Kim; FIGS. 4-316 and 318; col. 10, lines

5-36) taught by Kim in in the process of delineating the "three region (ROIs: G, R1, R2)"

(Choel; Fig. 2; Chapter Material and Methods: Ex vivo materials, page 365) of Choel's

method of "evaluate the potential use of dual energy x-ray absorptiometry for the

assessment of bone mineral content and bone mineral density" (Choel; Objective; page

364).

The suggestion/motivation for doing so would have been to provide "a contrast

enhancer based on mean-separate histogram equalization having ... brightness

compensation" (Kim; col. 4, lines 36-41) for "image quality enhancing" (Kim; col. 4, line

37), and especially, when the "mean-separated histogram equalization is applied, an

abrupt change in brightness and artifacts, which can be generated after a general

histogram equalization when an input image has a concentrated distributed histogram, can be effectively prevented" (Kim; col. 4, lines 46-51).

Therefore, it would have been obvious to combine Choel and Kim to obtain the invention as recited in claim 12.

b.   Regarding claim 13, the combination of Choel and Kim, as applied in claim 12, discloses all the previous claim limitations. Moreover, Kim discloses further comprising setting means for setting said standard value (FIGS. 5a and 5b; "[t]he corrected values (Δ) [being] determined by correction function" at col. 9, lines 43-47).

Choel and Kim are combinable because they are all related to the field of image processing.

At the time of the invention, it would have been obvious to a person of ordinary skill in the art to apply "[t]he corrected values (Δ) [being] determined by correction function" (Kim; FIGS. 5a and 5b; col. 9, lines 43-47) taught by Kim in the process of delineating the "three region (ROIs: G, R1, R2)" (Choel; Fig. 2; Chapter Material and Methods: Ex vivo materials, page 365) of the combination of Choel and Kim.

The suggestion/motivation for doing so would have been to provide "[a] brightness compensation [that] can be simply carried out by mapping a current mean to a desired out mean during the mean-separated histogram equalization" (Kim; col. 4, lines 57-59) when it is applied, "an abrupt change in brightness and artifacts, which can be generated after a general histogram equalization when an input image has a concentrated distributed histogram, can be effectively prevented" (Kim; col. 4, lines 46-51).

Therefore, it would have been obvious to combine Choel and Kim to obtain the invention as specified in claim 13.

c.   Regarding claim 14, the combination of Choel and Kim, as applied in claim 12, discloses wherein said standard average (Kim; equation (17); "a compensated mean level (B_m) $B_m = X_m + \Delta$" at col. 9, lines 34-35) and said standard deviation (Kim; equations (18) and (19); "$c_L(X_k)B_m$" and "$B'_m + (X_{L-1} - B'_m)c_U(X_k)$" at col. 10, lines 21-25) being set based on a result of detection by said detecting means (Kim; FIG. 4-304, "frame mean calculator; "A frame mean calculator calculates the mean level (X_m)" at col. 8, lines 38-39; FIGs. 4-308, "a first CDF calculator" and 4-310, "a second CDF calculator"; equations (15) and (16); "calculate[ing] a cumulative density function (CDF) $c_L(X_k)$ ... calculate[ing] a cumulative density function (CDF) $c_U(X_k)$" at col. 8, lines 63-67 to col. 9, lines 1-18) of a particular X-ray picture (Choel; Fig. 2; Chapter-Material and Methods: Ex vivo materials, page 365, "The author discloses that three regions of interest- G, R1 and R2 were delineated from the three specimens- Incisal specimen (IS), premolar specimen (PS) and molar specimen (MS), by the x-ray.").

d.   Regarding claim 16, the combination of Choel and Kim, as applied in claim 12, discloses wherein said evaluating means including judging means for judging said bone mineral density (Choel; Fig. 2; Table III; Chapter-BMD differences related to sex, dental status, and anatomic location, page 367, "The author discloses using a multiple regression analysis based on the mean bone mineral density (BMD) and SDs for G, R1 and R2 to assess the sex and dental status.") on the basis of said corrected gradation (Kim; equation (18); "equalized output (Yo)" at col. 10, lines 22-25 and 37-38).

e.   Regarding claim 17, the combination of Choel and Kim, as applied in claim

12, discloses further comprising output means for providing together a plurality of

evaluation results provided by said evaluation means (Choel; Fig. 2; Table III; Chapter-

BMD differences related to sex, dental status, and anatomic location, page 367, "The

author discloses using a multiple regression analysis based on the mean bone mineral

density (BMD) and SDs for G, R1 and R2 to assess the sex and dental status.") for

respective ones of a plurality of X-ray pictures (Choel; Fig. 2; Table III; Chapter-BMD

differences related to sex, dental status, and anatomic location, page 367, "G, R1 and

R2 obtained for male and female specimens, and dentate and edentulous specimens").

12.   Claim 15 is rejected under 35 U.S.C. 103(a) as being unpatentable over Choel et

al. ("Trabecular alveolar bone in the human mandible: A dual energy x-ray

absorptiometry study") in view of Kim (US 6,078,686), and further in view of Inoue (US

2001/0021269).

a.   Regarding claim 15, the combination of Choel and Kim, as applied in

claim 12, discloses all the previous claim limitation including wherein said evaluating

means (Choel; Table III; Chapter- Material and Methods: Ex vivo materials and Chatper-

BMD differences related to sex, dental status, and anatomic location, pages 365 and

367, "The author discloses using a multiple regression analysis based on the mean

bone mineral density (BMD) and SDs for G, R1 and R2, which are the global specimen,

infra-alveolar ROI and real intra-alveolar ROI, respectively, to assess the sex and dental

status.").

However, the combination of Choel and Kim does not disclose display means for displaying said corrected gradation in the form of histogram.

Instead of Choel and Kim, Inoue, the same field of endeavor of image processing, discloses display means (Fig. 3-310; "display" at ¶ 0118) for displaying said corrected gradation in the form of histogram (Fig. 3-307; "The author discloses that a histogram, which is created after gradation conversion has passed through the memory, is displayed on the display as the histogram" at ¶ 0118).

Choel, Kim and Inoue are combinable because they are all related to the field of image processing.

At the time of the invention, it would have been obvious to a person of ordinary skill in the art to apply the step of "[t]he created histogram [being] displayed on the display 310 as the histogram 307" (Inoue; Figs. 3-307 and 3-310; ¶ 0118) in the step of the "multiple regression analysis ... [which] sex and dental status [being] assessed' (Choel; Table III; Chapter- Material and Methods: Ex vivo materials and Chatper-BMD differences related to sex, dental status, and anatomic location, pages 365 and 367) of the combination of Choel and Kim.

The suggestion/motivation for doing so would have been to provide "[t]he created histogram [being] displayed on the display at the histogram" (Inoue; ¶ 0118) because

"an observer can observe most easily" (Inoue; ¶ 0067) "a histogram of an image ...

while gradationally converting one or a plurality of images interactively" (Inoue; ¶ 0067).

Therefore, it would have been obvious to combine Choel, Kim and Inoue to obtain the invention as specified in claim 15.

Conclusion

13.   No claims are allowed.

14.   The prior art made of record is considered pertinent to the disclosure of the application:

- Guillemaud (US 6,296,387 B1): The invention related to a method for correcting image defects from a matrix-type X or γ-ray detector, consisting in producing a confidence map.

- Lang et al. (US 2003/0112921 A1): The invention relates to a method and devices for analyzing x-ray images. In particular, devices , methods and algorithms are provided that allow for the accurate and reliable evaluation of bone structure from x-ray images.

15.   Any inquiry concerning this communication or earlier communications from the examiner should be directed to JOHN Wahnkyo LEE whose telephone number is (571)272-9554. The examiner can normally be reached on Monday - Friday (Alt.) 7:30 a.m. - 5:00 p.m..